



# NZYSpeedy qPCR Green Master Mix (2x), ROX plus

## Catalogue number:

MB22201, 2 mL (200 x 20 µL)  
MB22202, 5 mL (500 x 20 µL)  
MB22203, 20 mL (2000 x 20 µL)

## Description

NZYSpeedy qPCR Green Master Mix (2x), ROX plus is an optimized and highly efficient reaction mixture developed for real-time PCR. This master mix enables fast and highly reproducible procedures on the most common real-time PCR apparatus. The latest developments in PCR enhancers have been incorporated in this master mix, including buffer chemistry and a polymerase with hot start like activity. These combinations guarantee that NZYSpeedy qPCR Green Master Mix (2x), ROX plus delivers ultra-sensitivity coupled with highly reproducible and fast real-time PCR protocols. The master mix is provided as a 2x reaction mixture that contains all components necessary for real-time PCR, including a green intercalating dye, dNTPs, stabilisers and enhancers. NZYSpeedy qPCR Green Master Mix (2x), ROX plus is ready-to-use and only requires primers and template addition. It is optimized for intercalating green dye detection on different instruments.

## Storage temperature

NZYSpeedy qPCR Green Master Mix (2x), ROX plus should be stored at -20 °C in a constant temperature freezer. Minimize the number of freeze-thaw cycles by storing in working aliquots. The green dye is light sensitive, therefore the master mix should be protected from light whenever possible.

## Compatible real-time PCR instruments

The master mix is compatible with instruments that measure the passive reference signal. However, it is also compatible with instruments that do not require a passive reference signal for data normalization. The NZYSpeedy qPCR Green Master Mix (2x), ROX plus was optimized to be compatible with the following real-time PCR instruments:

Applied Biosystems™: 7000; 7300; 7700; 7900; 7900HT; 7900HT FAST; StepOne™ & StepOne™plus

## Protocol

The following protocol serves as a general guideline and a starting point for any qPCR procedure. Optimal reaction conditions (incubation times and temperatures, concentration of template DNA) may vary and, in particular conditions, may require further optimization.

**Reaction mix composition:** the given volumes are based on a standard 20 µL final reaction mix which can be scale adjusted.

Component	Volume	Concentration
NZYSpeedy qPCR Green Master Mix (2x), ROX plus	10 µL	1x
10 µM forward primer	0.8 µL	400 nM
10 µM reverse primer	0.8 µL	400 nM
Template	up to 8.4 µL	
Nuclease-free water	as required	

**Testing and Ct values:** When comparing NZYSpeedy qPCR Green Master Mix (2x), ROX plus with a mix from another supplier we strongly recommend amplifying from a 10-fold template dilution series. Loss of detection at low template concentration is the only direct measurement of sensitivity. An early Ct value is not an indication of good sensitivity, but rather an indication of speed. For difficult amplicons, increasing MgCl<sub>2</sub> concentration to 6 mM may reduce C<sub>t</sub>s.

## Suggested thermal cycling conditions

NZYSpeedy qPCR Green Master Mix (2x), ROX plus was optimized for the amplification of DNA fragments up to 200 bp under different real-time PCR cycling conditions. The table below displays a standard setup optimized on a number of platforms. However, these conditions may be adapted to suit different machine-specific protocols.

Cycles	Temp.	Time	Notes
1	*95 °C	*2 min	Polymerase activation
40	95 °C	5 s	Denaturation
	60-65 °C	**15-30s	Annealing/Extension (acquiring at end of step)

\*2 min for cDNA, up to 3 min for genomic DNA.

\*\*Recommendation: combined annealing/extension should be lower than 30 seconds.

## General considerations

In order to prevent any DNA contamination, we recommend that users have independent areas for reaction set-up, PCR amplification and any post-PCR gel analysis. It is essential that any tubes containing amplified PCR product are not opened in the PCR set-up area.

**Primers:** The specific amplification, yield and efficiency of any real-time PCR can be affected by both sequence and primers concentration, as well as by the fragment length. We strongly recommend taking the following suggestions into consideration when designing and running your real-time PCR experiment:

- Primers should have a melting temperature ( $T_m$ ) of approximately 60 °C.
- The fragment length should be between 80-200 bp and not superior to 400 bp
- Final primer concentration of 400 nM is suitable for most *green* reactions. However, to determine the optimal concentration we recommend titrating in the range 0.1-1 µM. The forward and reverse primers concentration should be equimolar.
- Design intron spanning primers when amplifying from cDNA (to avoid gDNA amplification).

**Template:** It is important that the DNA template is purified and concentrated according to conventional nucleic acid clean up procedures (NZYGelpure, MB011). In addition, templates must be devoid of any contaminating PCR inhibitors (e.g. EDTA). The recommended amount of template for PCR is dependent upon the type of DNA used. Please consider the following points when using genomic DNA or cDNA templates:

- **Genomic DNA:** use up to 1 µg of genomic DNA in a single PCR. We recommend using NZY Tissue gDNA Isolation kit (MB135) for high yield and purity from both prokaryotic and eukaryotic sources.
- **cDNA:** the optimal amount of cDNA to use in a single PCR is dependent upon the copy number of the target gene; we suggest using 100 ng cDNA per reaction. However, it may be necessary to vary this amount performing a two-step RT-PCR. We suggest using NZY First-Strand cDNA Synthesis Kit (MB125) for reverse transcription of purified RNA. To obtain high yield of highly purified RNA we suggest using the NZY Total RNA Isolation Kit (MB134).

**MgCl<sub>2</sub>:** NZYSpeedy qPCR Green Master Mix (2x), ROX plus contains MgCl<sub>2</sub> at a concentration of 3 mM, in the final 1x reaction mix, which is an optimal concentration for most real-time PCR procedures.

**PCR controls:** The reliability of the data may be affected by the presence of contaminating DNA, so it is important to detect it. We suggest that you always include a no-template control reaction, replacing the template with PCR-grade water. When performing a two-step RT-PCR, set up a no-RT control as well as a no-template control for the PCR. Furthermore, refer to the instrument instructions for the option of melt-profile analysis.

### Quality control assays

#### Genomic DNA contamination

NZY qPCR Master Mixes (2x) must be free of any detectable genomic DNA contamination as evaluated through real time PCR.

#### Nuclease assay

0.2-0.3 µg of pNZY28 plasmid DNA are incubated with NZY qPCR Master Mixes (2x) for 14-16 hours at 37 °C. Following incubation, the DNA is visualised on a GreenSafe-stained agarose gel. There must be no visible nicking or cutting of the nucleic acid.

#### Functional assay

NZY qPCR Master Mixes (2x) are extensively tested for activity, processivity, efficiency, sensitivity and heat activation.

V1901

### Certificate of Analysis

Test	Result
Genomic DNA contamination	Pass
Nuclease contamination	Pass
Functional assay	Pass

Approved by:



Patrícia Ponte  
Senior Manager, Quality Systems

For research use only.

